

**IEC****IECEE**  
CB  
SCHEME

Ref. Certif. No.

**UA-0187**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

**CB TEST CERTIFICATE**

Product

**Capacitor charging power supply**

Name and address of the applicant

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk, Belarus

Name and address of the manufacturer

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk, Belarus

Name and address of the factory

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk, Belarus

Note: When more than one factory, please report on page 2

 Additional Information on page 2

Ratings and principal characteristics

Input:200-240VAC, 50-60Hz, 24-19,5 A, Class I  
Output: 1000VDC

Trademark / Brand (if any)

OEM Tech

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

PCA-40-1000V-PD

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012 inclusive National differences for USA, Canada

As shown in the Test Report Ref. No. which forms part of this Certificate

1322-2-2020

This CB Test Certificate is issued by the National Certification Body

4, Metrologichna, str., 03143 Kyiv, Ukraine  
Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical Equipment  
UkrTEST of SE "UKRMETRTESTSTANDART"

Date: 2021-03-24

Signature

Victor Rymer



Ref. Certif. No

**UA-0192**

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product

**Capacitor charging power supply**

Name and address of the applicant

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk,  
Belarus

Name and address of the manufacturer

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk,  
Belarus

Name and address of the factory

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk,  
Belarus

Note: When more than one factory, please report on page 2

Additional Information on page 2

Ratings and principal characteristics

Input:200-240VAC, 50-60Hz, 24-19.5 A  
Output: 1000V

Trademark / Brand (if any)

**OEM Tech**

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

**PCA-40-1000V-PD**

Additional information (if necessary may also be reported on page 2)

The risk management requirements of the standard were not addressed

Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1-2:2014 inclusive National differences for Canada

As shown in the Test Report Ref. No. which forms part of this Certificate

2018-5-2020

This CB Test Certificate is issued by the National Certification Body

4, Metrologichna, str., 03143 Kyiv, Ukraine  
Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical Equipment  
UkrTEST of SE "UKRMETRTESTSTANDART"

Date: 2021-06-09

Signature:

Victor Rymer

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product

**Capacitor charging power supply**

Name and address of the applicant

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk,  
Belarus

Name and address of the manufacturer

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk,  
Belarus

Name and address of the factory

OOO "OEM Tech",  
Odoevskogo Str.129-106, 220018 Minsk,  
Belarus*Note: When more than one factory, please report on page 2* Additional information on page 2

Ratings and principal characteristics

Input:200-240VAC, 50-60Hz, 21-18 A, Class I  
Output: 1300VDC

Trademark / Brand (if any)

OEM Tech

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

PCA-40-1300V-CD

Additional information (if necessary may also be reported on page 2)

 Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012 inclusive National differences for USA, Canada

As shown in the Test Report Ref. No. which forms part of this Certificate

2019-2-2020

This CB Test Certificate is issued by the National Certification Body

4, Metrologichna, str., 03143 Kyiv, Ukraine  
Ukrainian Scientific and Technical Institute for Certification and Testing of Electrical  
Equipment UkrTEST of SE "UKRMETRTTESTSTANDART"

Date: 2021-02-24

Signature:

Victor Rymer



# OEM Tech

LASER ELECTRONICS

OEM Tech LLC. declares under its sole responsibility that the product(s) to which this declaration relates, is in conformity with the following directives, standard(s) and other normative document(s).

**List of models:** PCA-40-1000V-PD

**Directives:** Medical Device Directive  
93/42/EEC  
Electromagnetic Compatibility Directive  
2014/30/EU  
Restriction of Hazardous Substances Directive  
2011/65/EU

**Product Safety Standards:** IEC 60601-1:2005/AMD1:2012

**EMC Standards:** IEC 60601-1-2:2014

Note: This product is classified as component type for building-in use. The EMC characteristic and safety approval of the final application has to be performed by the end-product manufacturer in accordance to the applicable standards for the end-product.

June 01, 2021

  
Dmitry Kamlyuk, CEO



OEM Tech LLC  
129 Odoevskogo str.  
220018 Minsk  
Belarus

T. +375 17 3224054  
F. +375 17 3224064  
[info@oem-tech.by](mailto:info@oem-tech.by)  
[www.oemtex.com](http://www.oemtex.com)

